
DATE	REV	REVISION RECORD	AUTH.	DR.	CK.



SEAPOINT	GAIN	RANGE
TURBIDITY METER	5X	500 FTU
FLUOROMETER	3X	50 ug/l

CABLE	P1 AG-206	P2 AG-206	P3 RMG-4FS	P4 VMG-4FS
COMMON	PIN 1	PIN 1,3,5	PIN 1,3	PIN 1,4
OXYGEN	PIN 2	--	PIN 2	--
SPARE	PIN 3	--	--	--
SEAPOINT	PIN 4	PIN 2	--	--
WETLABS	PIN 5	--	--	PIN 2
POWER	PIN 6	PIN 4	PIN 4	PIN 3

SBE PART NO.	DIM A	DIM B	DIM C	DIM D
172027	24 INCHES	28 INCHES	28 INCHES	28 INCHES

TOLERANCES	SEA-BIRD ELECTRONICS, INC		
FRACTIONAL	P/N SEE TABLE	SCALE NTS	DRAWN BY KHH APPROVED BY
DECIMAL	TITLE Y - CABLE: SBE43/SEAPOINT/WETLABS		
ANGULAR	DATE 11.9.2004	DRAWING NUMBER 32991	REV -